

Table 38: Device Mastery by Step 1 of Stage 1

Device mastery by step 1	Empty SPIROMAX	Empty TURBOHALER	Odds ratio (95% CI)^a	p-value^b
Full analysis set (FAS), N patients	481	481		
Yes	160 (33.3%)	55 (11.4%)	4.89 (3.23, 7.40)	<0.001
No	321 (66.7%)	426 (88.6%)		

a Empty SPIROMAX relative to empty TURBOHALER (Empty TURBOHALER=1.00). The statistical method was conditional logistic regression.

b p-value as chi-square; p<0.05 considered statistically significant .

CI=confidence interval; N=number.