

Table 46: Patient Preference for the Device in Stage 1

Total Patient Satisfaction and Preference Questionnaire score	Empty SPIROMAX	Empty TURBOHALER	Median difference (IQR)	p-value^a
Full analysis set, N patients	477	477		
Median (min, max)	89.80 (18.37, 100.00)	85.71 (14.29, 100.00)	2.04 (10.20)	<0.001

a p-value as Wilcoxon signed rank test; p<0.05 considered statistically significant.

CI=confidence interval; min=minimum; max=maximum; IQR=interquartile range; N=number.