

Table 24: Stage 1 Device Mastery

Device mastery	Empty SPIROMAX	Empty TURBOHALER	p-value ^a	Odds ratio (95% CI) ^b
Full analysis set(FAS), N patients	481	481		
Yes	454 (94.4%)	418 (86.9%)	<0.001	3.77 (2.05, 6.95)
No	27 (5.6%)	63 (13.1%)		

a The p-value for the treatment comparison is based on chi-square. $p < 0.05$ considered statistically significant.

b Empty SPIROMAX relative to empty TURBOHALER. Conditional logistic regression was used. CI=confidence interval; N=number.